

Operating Instructions

GANN HYDROMETTE COMPACT B

General Information

The Hydromette Compact B is an electronic moisture indicator based on the dielectric-constant / high-frequency measurement principle.

The device is used for non-destructively tracing dampness in building materials of all kinds as well as for detecting damp distribution in walls, ceilings and floors.

The device is particularly suitable for pre-testing the readiness of building materials for covering prior to CM measurement.

Measuring Range: **0 – 100 digits**

The magnitude of the value measured is mainly determined by the raw density and the water content of the material. The depth of penetration of the measuring field, depending on raw density, is approx. 20 –40 mm.

Adjustment

The device is calibrated automatically. Readjustment is not necessary.

Battery

9V block battery. Type IEC 6 F 22 or IEC 6 LF 22.

Changing the Battery

The battery requires changing when two decimal points are shown in the display (e.g. 1.8.8). Loosen the two cross-headed screws on the upper side of the device and carefully remove the cover in an upward direction. Change the battery and refit the cover.

Safety Advice

There is a **danger of injury** if the metal ball comes into contact with live parts. Do not use the device in the immediate vicinity of older equipment or equipment equally sensitive to high frequencies (e.g. functioning medicinal devices). Use the device **only** for measuring the dampness in hardened building materials by bringing the ball into contact with the surface.

Check

Hold the device as near to the lower part of the housing, opposite to the ball sensor, as possible. Press the start button and hold the device with the ball in the air. The displayed value must vary between –5 and +5.

Operation

Hold the device as near to the lower part of the housing, opposite to the ball sensor, as possible. Press the measuring button and use the ball to scan the surface under investigation. The ball must be in firm contact with the material. To obtain the best results, the device should be held at an angle of 90° to the surface to be measured.

Please note:

Do not take measurements on metallic linings!

In corners or recessed areas a distance of approx. 8 – 10 cm from the edge / recess must be maintained.

If there is any metal in the substructure (constructional steel, channels, pipes, plaster rails, etc.) and with normal coverings, the display jumps to approx. 50 digits for otherwise dry surroundings.

It is only possible to reach a conclusion about the absolute dampness in wt.-% or the dampness in CM-% if the normal drying process has taken place (e.g. not during or shortly after the use of drying agents or heat guns). If there is not a roughly normal variation in dampness between the surface and the interior, too low a measured value may be indicated.

The raw density of the material being measured has a noticeable affect. Basically, the value displayed with dry and damp building materials increases correspondingly with increasing raw density.

The values given in the table below are indicative and non-binding. Please bear in mind, when evaluating the measured value displayed on the GANN HYDROMETTE COMPACT B with respect to the material, that it is not a dampness measurement qualified to VOB or the relevant specialist regulations.

All information and tables in these operating instructions concerning permissible or common dampness conditions in practice as well as the general definition of terms are taken from technical literature. The manufacturer of the device can thus not give any guarantee for the correctness of this information. The conclusions to be drawn from the results of measurements by each user depend upon the individual circumstances and his experience gained from professional practice.

This publication replaces all previous versions. It may not be reproduced or electronically processed, duplicated or distributed in any way without the written permission of Gann Mess- u. Regeltechnik GmbH. Subject to technical and documentary change. All rights reserved. This document was compiled with the necessary care. Gann Mess- u. Regeltechnik GmbH accepts no liability whatsoever for errors or omissions.

Table of Comparison

Structural Moisture

LCD - Readout	COMPACT B	0	10	20	30	40	50	60	70	80	90	100
Gypsum + anhydride based constr.mat. / flooring - Wt. % / CM %					- 0,3 -	- 0,5 -	- 1,4 -	- 2,0 -	- 2,3 -	- 2,7 -		%
Cement based construction materials and flooring in CM %						- 1,5 -	- 2,1 -	- 3,0 -	- 3,5 -	- 4,0 -		CM-%
Cement based construction materials and flooring - in weight %						- 2,7 -	- 3,6 -	- 4,5 -	- 5,5 -	- 6,0 -		Wt. - %

CM % = percentage value by carbide method

© Copyright by Gann Mess- u. Regeltechnik GmbH

07/2016

WEEE Directive 2002/96/EC Electrical and Electronic Equipment Act

The packaging, battery and instrument must be disposed of at a recycling centre in accordance with legal regulations.
The instrument was manufactured after 01.10.2009